Stereospecific Detection Technologies
***Reagents beyond the chemistry***

SDT GmbH • Pascalstraße 17
52499 Baesweiler • Germany
www.sdt-reagents.de

## ***Certificate of Analysis***

| | | | |
|---|---|---|---|
| Product: | **Streptavidin-PolyHRP10** | | |
| Code #: | **#SP10C** | | |
| Concentration: | 1 mg/ml **total conjugate** | | |
| Buffer base: | PBS pH 7,3 + 50% ultra-pure glycerol | | |
| Stabilizing additives: | HRP stabilizer (proprietary)<br>Ultra pure biotin-free BSA | -<br>- | 100 µg/ml<br>15 mg/ml |
| Preservative/anti-microbial: | 5-Bromo-5-nitro-1,3-dioxane | - | 1500 ppm |
| Lot: | 196921 | | |
| Storage: | -20°C (-15°C/-25°C). Do not store below -25°C! | | |
| Expiration date: | 04/2029 | | |
| Quality Control Assurance Release Procedure: | ELISA use test with hIgG-biotin as analyte at different conditions | | |

Result: POSITIVE (COMPLIES) ~~NEGATIVE (DOES NOT COMPLY)~~

Date: 12.03.2026

This document has been electronically produced and is valid without a signature