Stereospecific Detection Technologies
Reagents beyond the chemistry

SDT GmbH • Pascalstraße 17
52499 Baesweiler • Germany
www.sdt-reagents.de

Certificate of Analysis

Product:	**Streptavidin-PolyHRP20**
Code #:	#**SP20D50-RHT**
Concentration:	50 µg/ml pre-diluted **SA-PolyHRP20**, liquid formulation stabilized for storage at +4°C (format 1/20)
Buffer base:	Streptavidin-PolyHRP Stabilizer RHT, #**SA3-RHT**
Lot:	766586
Storage:	+2°C/+8°C. Do not freeze below 0°C!
Expiration date:	03/2028
Quality Control Assurance Release Procedure:	ELISA use test with hIgG-biotin as analyte at different conditions
Result:	POSITIVE (COMPLIES) — ~~NEGATIVE (DOES NOT COMPLY)~~
Date:	24.02.2026

This document has been electronically produced and is valid without a signature