



**Certificate of Analysis**

|                                                   |                                                                                                                |                                           |
|---------------------------------------------------|----------------------------------------------------------------------------------------------------------------|-------------------------------------------|
| Product:                                          | <b>monoclonal Mouse anti-Human IgA (H, <math>\alpha</math>)<br/>StrongZyme™ HRP Conjugate</b>                  |                                           |
| Code:                                             | <b>#MAHA-SZ</b>                                                                                                |                                           |
| Presentation:                                     | 2 mg/ml in PolyHRP Stabilizer SA1                                                                              |                                           |
| Lot:                                              | 872621                                                                                                         |                                           |
| Storage:                                          | +2°C ~ +8°C. Do not freeze!                                                                                    |                                           |
| Expiration date:                                  | 04/2022                                                                                                        |                                           |
| Quality Control Assurance<br>Release Procedure:   | ELISA use test with adsorbed hIgA, hIgG, hIgM and hIgE<br>proving no cross-reactivity with hIgG, hIgM and hIgE |                                           |
| Recommended working strength<br>dilution in ELISA | 1:40 000 – 1:50 000<br>(e.g. Chlamidia IgA)                                                                    |                                           |
| Result:                                           | <div>POSITIVE<br/>(COMPLIES)</div>                                                                             | <div>NEGATIVE<br/>(DOES NOT COMPLY)</div> |
| Date:                                             | 18.03.2020                                                                                                     |                                           |

This document has been electronically produced and is valid without a signature